# Micron III/IV EQUIPMENT PROCEDURES

Failure to follow these procedures will result in loss of access to the Micron IV equipment.

### 1. Access Policy

- 1) All users must be trained before using the Micron IV equipment. Contact Systems Analysts to schedule training sessions.
- 2) Enter your name every time when you use the Micron IV equipment.
- 3) Do not allow anyone to work under your name.

#### 2. Availability

- 1) Micron IV equipment is available 24 hours/day and 7 days/week.
- 2) Sign up using Google Calendar during working hours at least 48 hours in advance of your session.
- 3) There are four six-hour sessions available per day:

 Session 1.
 7:00 a.m. – 1:00 p.m.

 Session 2.
 1:00 p.m. – 7:00 p.m.

 Session 3.
 7:00 p.m. – 1:00 a.m.

 Session 4.
 1:00 a.m. – 7:00 a.m.

- 4) If you need more than one 6-hour session/ day, you must have permission from the Systems Analysts.
- 5) If you need less than 6 hours, indicate your actual time of use on the Google Calendar.
- 6) If you must cancel your Micron IV session, you must delete your sign-up on the Google Calendar.
- 7) Those who consistently sign up but do not use their reserved times will lose access to the Micron IV equipment.
- 8) Users will be notified of maintenance downtime at least one week prior to scheduled maintenance.

### 3. Micron III Equipment Use

- 1) Change to clean cages each time you bring animals to the Micron IV room.
- 2) Do not move the Micron IV equipment out of the Micron IV room.
- 3) Do not change any computer settings.
- 4) Do not attempt to repair the equipment if it is not functioning. Contact System Analysts immediately.
- 5) If reagent supplies are low, contact Systems Analysts.
- 6) Fill out the Micron IV Use Log sheet at the end of each session.
- 7) Clean the work area when your session is complete.

## 4. Systems Analysts Contact Information

Feng Li: <u>feng-li@ouhsc.edu</u> 42820 or 229-8717

Nanette Wheatley: nanette-wheatley@ouhsc.edu 361-7473

Failure to follow these procedures will result in loss of access to the Micron IV equipment.

DIRECTOR	 	 
DVM		